



ORIGINAL ARTICLE

Predictive capacity of different bioelectrical impedance analysis devices, with and without protocol, in the evaluation of adolescents[☆]

Vivian Siqueira Santos Gonçalves^{a,*}, Eliane Rodrigues de Faria^b,
Sylvia do Carmo Castro Franceschini^a, Silvia Eloiza Priore^a

^a Department of Nutrition and Health, Universidade Federal de Viçosa, Viçosa, MG, Brazil

^b Universidade Federal do Espírito Santo, Alegre, ES, Brazil

Received 22 October 2012; accepted 13 March 2013

Available online 13 September 2013

KEYWORDS

Adolescent health;
Body composition;
Electric impedance;
Obesity

Abstract

Objective: this study was performed to determine the predictive capacity of four different bioelectrical impedance analysis (BIA) devices in the assessment of adolescents, with and without a protocol.

Methods: a cross-sectional study was performed with 215 adolescents aged 10 to 14 years, of both genders, evaluated through anthropometry and body composition by dual energy X-ray absorptiometry (DXA) and by four different BIA devices, with and without a protocol. The following tests were used: Kolmogorov-Smirnov's, chi-squared, Student's *t* or Mann-Whitney's, Kruskal-Wallis's, Wilcoxon's, and kappa index. The ROC curves were constructed and the sensitivity, specificity, and positive and negative predictive values were calculated.

Results: of the 215 adolescents, 44.2% had excessive body fat. The tetrapolar BIA device equipped with eight tactile electrodes showed more sensitivity and results that were closer to those obtained by DXA (area under the ROC curve [AUC] = 0.964 with protocol and AUC = 0.973 without protocol, $p < 0.001$), as well as greater agreement ($k = 0.67$ with protocol and $k = 0.63$ without protocol, $p < 0.001$). The evaluation without protocol was similar to that by DXA in most investigated situations ($p > 0.05$).

[☆] Please cite this article as: Gonçalves VS, Faria ER, Franceschini SC, Priore SE. Predictive capacity of different bioelectrical impedance analysis devices, with and without protocol, in the evaluation of adolescents. J Pediatr (Rio J). 2013;89:567–74.

* Corresponding author.

E-mail: vivian.goncalves@ufv.br (V.S.S. Gonçalves).

PALAVRAS-CHAVE

Saúde do adolescente;
Composição corporal;
Impedância elétrica;
Obesidade

Conclusion: BIA is capable of predicting alterations in adolescents' body composition. When it is impossible to perform the assessment with a protocol, its results may be useful in population studies.

© 2013 Sociedade Brasileira de Pediatria. Published by Elsevier Editora Ltda.

Este é um artigo Open Access sob a licença de [CC BY-NC-ND](#)

Capacidade preditiva de diferentes equipamentos de bioimpedância elétrica, com e sem preparo prévio, na avaliação de adolescentes

Resumo

Objetivo: determinar a capacidade preditiva de quatro equipamentos distintos de bioimpedância elétrica (BIA) na avaliação de adolescentes, com e sem a realização de protocolo.

Métodos: estudo transversal realizado com 215 adolescentes de 10 a 14 anos, de ambos os sexos, avaliados através da antropometria e da composição corporal pelo DEXA e por quatro equipamentos distintos de BIA, com e sem protocolo. Foram utilizados os testes estatísticos: *Kolmogorov-Smirnov*, do Qui-quadrado, *t-Student* ou *Mann-Whitney*, *Kruskal-Wallis*, *Wilcoxon* e Índice Kappa. Foram construídas curvas ROC e calculados os valores de sensibilidade, especificidade e preditivos positivo e negativo.

Resultados: dos adolescentes, 44,2% apresentaram excesso de gordura corporal. A BIA tetrapolar, equipada com oito eletrodos táteis, demonstrou-se mais sensível e com resultados mais próximos ao DEXA (AUC = 0,964 com protocolo e AUC = 0,973 sem protocolo, $p < 0,001$), apresentando, também, maior concordância ($k = 0,67$ com protocolo, e $k = 0,63$ sem protocolo, $p < 0,001$). A avaliação sem protocolo foi semelhante ao DEXA na maioria das situações investigadas ($p > 0,05$).

Conclusão: a BIA é um instrumento capaz de prever distrofias relacionadas à gordura corporal de adolescentes. Na impossibilidade de realização do protocolo, seus resultados podem ser úteis em estudos populacionais.

© 2013 Sociedade Brasileira de Pediatria. Publicado por Elsevier Editora Ltda.

Este é um artigo Open Access sob a licença de [CC BY-NC-ND](#)

Introduction

The World Health Organization (WHO) defines adolescence as the period from 10 to 19 years, characterized by intense physical, psychological, and social changes. Rapid growth and nutritional vulnerability also characterize this phase, when there is consolidation of eating habits, which, when adequate, can become a protective factor for obesity, cardiovascular disease, and metabolic disorders in adulthood.¹

According to the household budget survey (HBS) conducted by the Brazilian Institute of Geography and Statistics (Instituto Brasileiro de Geografia e Estatística - IBGE) in the years 2008 and 2009, excess body weight was identified in approximately 20% of the adolescent population of the metropolitan areas of Brazil.²

When weight increases due to excess body fat, it can lead to adolescent obesity, which has been considered a predictor of risk for cardiovascular disease, diabetes mellitus, dyslipidemia, and hypertension. Therefore, it is relevant to adequately estimate body composition in adolescence, as it is a period of great change, mainly as a result of puberty.³⁻⁶

It must be emphasized, however, that there is a need for methods used in body composition determination that are practical, fast, and easy to perform, with the possibility of being applicable to several working conditions, including in population-based studies in the field. Among these methods, bioelectrical impedance analysis (BIA) is highlighted, as it has all these characteristics at a relatively low cost, in addition to its portability and noninvasiveness.^{4,5,7-9}

The use of previous preparations (protocols) for standardization of variables that affect body hydration is a recommendation to perform BIA.⁹⁻¹¹ However, its use may be restricted by lack of adherence or difficulty to follow these requirements by the adolescent.

Given the importance of accurately determining body composition and the broad use of BIA, this study aimed to determine the predictive capacity of four different devices in the evaluation of adolescents with and without a protocol.

Methods

Sample

This was an epidemiological, cross-sectional study, with a population of 215 adolescents of both genders, aged between 10 years to 14 years, 11 months, selected by simple random sampling from all public and private schools in the age range of interest, located in urban and rural areas of the city of Viçosa, state of Minas Gerais, Brazil. The following inclusion criteria were used: interest in participating in the study; absence of prosthetics and/or pacemakers; absence of chronic diseases or use of continuous medication that could interfere with body hydration; and adherence to the recommended protocol to undergo BIA.

Sample selection was based on the total number of adolescents in the city at the age of interest in 2010.¹² The sample was calculated using EpiInfo software, release 6.04 for cross-sectional studies, considering a total population of

5,754 individuals, the expected frequency of excess body fat of 17.5%,¹³ and variability of 5%, totaling 214 individuals, with a confidence level of 95%.

The sample draw was conducted among all who met the inclusion criteria and returned the signed informed consent, respecting the proportionality of the number of students that each school had in each age group. When the adolescent did not want to participate or abandoned the study, a new draw was made to replace him/her.

The project was approved by the Ethics Committee on Human Research of Universidade Federal de Viçosa (protocol. N. 0140/2010); adolescents and their parents signed the consent form, prepared in accordance with standards established by 196/96 Resolution of the National Health Council.

Anthropometric assessment

Weight was measured on a digital scale with a maximum capacity of 150 kg and a sensitivity of 50 g, whereas height was measured using a portable stadiometer with an extension of 2.13 m and 0.1 cm resolution. Measurements were made in duplicate, allowing the use of the mean values between the two measurements. In cases where the difference exceeded 0.5 cm, new measurements were performed. The body mass index (BMI)/age and height/age indices were calculated to characterize the population, using as reference the cutoffs (Z-scores), established by the WHO.¹⁴

Body composition

The body fat percentage (BF%) was measured by a DXA equipment (Lunar Prodigy Advance DXA System - analysis version: 13,31, GE Healthcare) and estimated by the following BIA equipment: tetrapolar horizontal Biodynamics®, model 450 (BIA 1), tetrapolar vertical Tanita®, model BC-558 (BIA 2), tetrapolar vertical Biospace®, equipped with eight tactile electrodes, InBody® 230 model (BIA 3), and bipolar vertical Tanita®, model 2220 (BIA 4). BF% was analyzed according to the classification proposed by Lohman¹⁵, considering as excess body fat values $\geq 20\%$ for males and $\geq 25\%$ for females, and as low body fat percentage $< 10\%$ for males and $< 15\%$ for females. Assessment by BIA was performed in two stages, following the protocol proposed by Barbosa¹⁶ and later, within an average time period of 9±4 days, without the protocol.

The protocol involved previous preparation aiming to standardize the hydration status to undergo the BIA assessment and consisted of the following: be at least seven days after the last menstrual period and seven days before the next; undergo complete fasting in the previous 12 hours; refrain from physical exercises in the previous 12 hours; no alcohol consumption in the previous 48 hours; no use of diuretics for at least seven days before the assessment; and urination 30 minutes before the assessment. Adolescents were also asked to remove metal objects such as earrings, rings, watches, and others, which could interfere with the passage of electrical current.

Statistical analysis

The Kolmogorov-Smirnov normality test was performed to determine variable distribution (parametric or not) and thus

choose the most appropriate statistical test to evaluate data. Parameters with normal distribution were expressed as mean and standard deviation; those with non-normal distribution were expressed as median and range.

The chi-squared test was used to compare prevalence. In case of continuous variables, Student's *t* or Mann-Whitney's tests were performed to compare two, and Kruskal-Wallis's test was used to compare three or more; Wilcoxon's test was used when two evaluations were performed by the same individuals using the same equipment (with and without protocol).

The kappa index was used to determine the agreement between the assessments by BIA and DXA, classifying it according to the criteria by Landis and Koch¹⁷, with the following concordances: from 0 to 0.19: poor; 0.2 to 0.39: weak; 0.4 - 0.59: moderate; 0.6 to 0.79: substantial; 0.8 to 1.0: almost perfect.

ROC curves were constructed to verify the capacity of BIA in predicting excess body fat when compared to DXA. Areas under the curve (AUC) were calculated with their respective 95% confidence intervals. The null hypothesis would be accepted with an AUC value ≤ 0.50 . Sensitivity (probability of a test being positive, if there is an alteration), specificity (probability of a test being negative, if there is no alteration), positive predictive values (proportion of true positives among all individuals who tested positive), and negative predictive values (proportion of true negatives among all individuals who tested negative) were calculated for each device, with and without protocol; excess body fat was considered as the altered variable.¹⁸

The database was created using Microsoft Office Excel 2007 with duplicate entries; statistical analyses were performed using the SIGMA STAT software, release 3.2 and MEDCALC statistical software, release 12.2.1.0. Significance level was set at $p < 0.05$.

Results

Subjects' characteristics

A total of 215 adolescents (of whom 53.5%, $n = 115$, were females), participated in the study, presenting the following median values for age, weight, and BMI: 11.9 years (range: 10.1 to 14.9), 42.2 kg (range: 25.1 to 92.8), and 18.0 kg/m² (range: 12.5 to 33.8). The mean height was 151.6±10.0 cm. There was no difference regarding these parameters between genders ($p > 0.05$).

Regarding the nutritional status, 2.8% ($n = 6$) had short stature for age, and 3.3% ($n = 7$) had low BMI for age, 16.7% ($n = 36$) were overweight, and 8.4% ($n = 18$) were obese.

DXA assessment showed a prevalence of 44.2% ($n = 95$) of excess BF% and 13.5% ($n = 29$) of low BF%.

Comparison of assessments with and without protocol

Table 1 shows the prevalence of low BF%, normal BF%, and high BF% measured by DXA and estimated by BIA, with and without protocol. It was observed that the evaluation carried out by all BIA devices with a protocol identified more adolescents with high BF% than without protocol.

Table 1 Nutritional status by percentage of body fat in adolescents assessed by DXA and four bioelectrical impedance devices, with and without protocol.

Device	Nutritional status by percentage of body fat					
	Low BF%		Normal BF%		High BF%	
	%	95% CI	%	95% CI	%	95% CI
DXA	13.5	9.6-18.7	42.3	35.9-49.0	44.2	37.7-50.1
BIA 1						
With protocol	6.5 ^a	3.9-10.6	55.8 ^a	49.1-62.3	37.7	31.5-44.3
Without protocol	9.3	6.1-13.9	54.0 ^a	47.3-60.5	36.7	30.6-43.4
BIA 2						
With protocol	0.0 ^a	0.0-1.8	46.5	40.0-53.2	53.5	46.8-60.0
Without protocol	0.0 ^a	0.0-1.8	53.5 ^a	46.8-60.0	46.5	40.0-53.2
BIA 3						
With Protocol	9.3	6.1-13.9	50.2	43.6-56.9	40.5	34.1-47.1
Without protocol	8.8	5.7-13.4	52.6 ^a	45.9-59.1	38.6	32.4-45.3
BIA 4						
With Protocol	6.5 ^a	3.9-10.6	59.5 ^a	52.9-65.9	34.0 ^a	28.0-40.5
Without protocol	10.7	7.2-15.5	56.3 ^a	49.6-62.7	33.0 ^a	27.1-39.6

BIA 1, biodynamics horizontal tetrapolar device, model 450; BIA 2, tanita vertical tetrapolar device, model BC-558; BIA 3, biospace equipped with eight tactile electrodes, model *InBody* 230; BIA 4, Tanita vertical bipolar device, model 2220; BF%, percentage of body fat; CI, confidence interval; DXA, dual-energy X-ray absorptiometry.

Chi-squared test: BIA vs. DXA.

^a $p < 0.05$.

Regarding the increase in BF%, BIA 4 was the only device that underestimated the prevalence ($p < 0.05$), whereas the others were similar to DXA in both assessments ($p > 0.05$). It is noteworthy that BIA 3 showed prevalence more similar to DXA in all situations ($p > 0.05$) except for normal BF% without protocol, where it overestimated it ($p < 0.05$).

When compared, all BIA devices had similar values for body fat in kg (BF) when compared to DXA in both with and without protocol assessments ($p > 0.05$), considering the total population. Regarding the stratification by gender, only the male gender, assessed by BIA 2, was higher than DXA ($p = 0.011$ with protocol and $p = 0.017$ without protocol). The comparison of BIA devices by gender is shown in [Table 2](#).

It was observed that, for females, the protocol did not influence any of the assessments, whereas for males, BIA 2 and 3 also showed similar values in both situations ($p > 0.5$). It is noteworthy, however, that BIA 3 did not differ from DXA, contrary to that occurred with BIA 2, which overestimated BF in both situations ($p < 0.05$).

Agreement between the assessments by BIA and by DXA

When analyzing the agreement between DXA evaluations and by each of the BIA devices ([Table 3](#)), significance were observed among all of them ($p < 0.001$). However, BIA 3 again

Table 2 Body fat determined by different electrical bioimpedance devices, with and without protocol, by gender, in adolescents.

BF (kg)	Female (n = 115)		p	Male (n = 100)		p
	With protocol Md (Range)	Without protocol Md (Range)		With protocol Md (Range)	Without protocol Md (Range)	
BIA 1	9.4 (3.2-35.0)	9.6 (2.4-23.6)	0.385	6.9 (2.4-28.3)	6.9 (2.2-28.2)	0.027
BIA 2	10.8 (5.6-26.7)	11.1 (5.5-24.6)	0.173	7.4 (2.9-31.6)	7.1 (3.2-30.8)	0.259
BIA 3	9.5 (2.7-30.4)	9.6 (3.0-30.7)	0.252	6.7 (1.6-33.8)	6.7 (1.3-35.1)	0.599
BIA 4	9.8 (2.3-27.8)	9.9 (2.4-28.0)	0.512	6.2 (1.0-31.8)	5.7 (1.9-30)	< 0.001

BIA 1, biodynamics horizontal tetrapolar device, model 450; BIA 2, tanita vertical tetrapolar device, model BC-558; BIA 3, biospace equipped with eight tactile electrodes, model *InBody* 230; BIA 4, tanita vertical bipolar device, model 2220; BF%, body fat; Md, median. Wilcoxon test (BIA with vs. BIA without protocol).

Table 3 Agreement between body fat assessment in adolescents performed by different bioelectrical impedance devices and DXA.

	Kappa index (n = 215)
BIA 1	
<i>With protocol</i>	
Total	0.52 ^a
Female	0.52 ^a
Male	0.50 ^a
<i>Without protocol</i>	
Total	0.59 ^a
Female	0.54 ^a
Male	0.64 ^a
BIA 2	
<i>With protocol</i>	
Total	0.49 ^a
Female	0.38 ^a
Male	0.57 ^a
<i>Without protocol</i>	
Total	0.51 ^a
Female	0.46 ^a
Male	0.53 ^a
BIA 3	
<i>With protocol</i>	
Total	0.67 ^a
Female	0.62 ^a
Male	0.70 ^a
<i>Without protocol</i>	
Total	0.62 ^a
Female	0.58 ^a
Male	0.65 ^a
BIA 4	
<i>With protocol</i>	
Total	0.53 ^a
Female	0.46 ^a
Male	0.62 ^a
<i>Without protocol</i>	
Total	0.52 ^a
Female	0.43 ^a
Male	0.61 ^a

BIA 1, biodynamics horizontal tetrapolar device, model 450; BIA 2, tanita vertical tetrapolar device, model BC-558; BIA 3, biospace equipped with eight tactile electrodes, model InBody 230; BIA 4, tanita vertical bipolar device, model 2220; Kappa index: BIA vs. DXA.

^a p < 0.001.

showed better results, with a strong agreement for the two assessments in both genders.

The other BIA devices also showed strong agreement for the male gender. For the total population and the female gender, agreement was considered moderate.

BIA 2 presented the worst results; the female gender, with protocol, was considered the weakest among all analyzed data.

Predictive capacity of electrical bioimpedance, with and without protocol

Table 4 shows the AUC, sensitivity, and specificity, as well as positive and negative predictive values for each device and gender, at the assessments with and without protocol, obtained after creating the ROC curves, considering the excess BF. It was observed that BIA 3, without protocol, showed the highest areas for the total population and for both genders, after stratification.

The AUCs were also compared for evaluations with and without protocol, performed for each gender and for each device and no differences were observed (p > 0.05).

Discussion

The nutritional status of the adolescents studied followed the trend indicated by the HBS, with a low frequency of malnutrition and higher prevalence of overweight. Approximately 25% of the adolescents in the city were overweight, higher than the national prevalence (20.5%) and that found in another study with adolescents from the state of Minas Gerais (20.1%), but within the range found in the South-eastern region of Brazil (20% to 27%).^{2,19} Comparing the prevalence of alterations in BF% (43%) and BMI/age (25%), it is clear that the index failed to identify several adolescents who already had these alterations, confirming the importance of methods to predict excess BF even in those who present normal weight.

It is worth mentioning that the BMI criteria should not be used alone. Adolescents with an adequate BMI may have a high BF% and may eventually have risks of morbidity similar to those with high BMI, especially in females²⁰⁻²⁴, highlighting the need for BF% assessment in order to identify possible risk factors for health.

Excess BF may be related to genetic, metabolic, physiological, and lifestyle components, such as sedentary lifestyle and poor eating habits. It is associated with insulin resistance, dyslipidemia, and metabolic syndrome, some of the risk factors for cardiovascular disease that have already been identified in adolescents at the age group analyzed in this study, corroborating the importance of monitoring these young individuals.^{20,21,23,25,26}

BIA devices were able to predict increases in BF%, but showed distinct characteristics when the protocol influence was analyzed. In general, when compared with DXA, it was observed that both assessments behaved similarly in relation to the prevalence of excess BF, and only BIA 4 showed a different result (p < 0.05). BIA 3 was shown to be the most stable device, differing from DXA in only one situation (normal BF without protocol), but similar to DXA regarding the prevalence of excess or low BF%, with and without the protocol. Therefore, for a prevalence study, BIA 3 appears to be the most adequate device, while BIA 4 is the least recommended.

BIA 1, 3, and 4 were similar to DXA when comparing BF in kg for both genders at the two assessments, indicating they are adequate for use. BIA 2, when evaluating the female gender, could also be considered useful. These devices are portable, easy to use and transport, and despite the price variation among them, are much more affordable than DXA.

Table 4 Predictive capacity of excess body fat in different electrical bioimpedance devices compared to DXA, by gender, in the evaluation of adolescents, 2011.

	AUC	SD	95% CI	Sensit%	Specif%	PPV%	NPV%
BIA 1							
<i>With protocol</i>							
Total	0.938 ^a	0.0144	0.897-0.967	91.6	79.2	77.7	92.2
Female	0.926 ^a	0.0220	0.862-0.966	72.1	96.3	95.7	75.4
Male	0.957 ^a	0.0173	0.897-0.987	91.2	87.9	79.5	95.1
<i>Without protocol</i>							
Total	0.946 ^a	0.0142	0.907-0.972	85.3	92.5	90.0	88.8
Female	0.931 ^a	0.0226	0.869-0.970	91.8	83.3	86.2	90.0
Male	0.966 ^a	0.0172	0.909-0.992	91.2	92.4	86.1	95.3
BIA 2							
<i>With protocol</i>							
Total	0.916 ^a	0.018	0.871-0.950	89.5	78.3	76.6	90.4
Female	0.919 ^a	0.0239	0.853-0.962	75.4	94.4	93.9	77.3
Male	0.980 ^a	0.0101	0.929-0.997	100.0	86.4	79.1	100.0
<i>Without protocol</i>							
Total	0.894 ^a	0.0215	0.845-0.932	80.0	85.3	81.7	84.4
Female	0.895 ^a	0.0286	0.824-0.944	88.5	77.8	81.8	85.7
Male	0.945 ^a	0.0266	0.880-0.981	85.3	92.4	85.3	92.4
BIA 3							
<i>With protocol</i>							
Total	0.964 ^a	0.0129	0.929-0.984	96.8	86.7	85.2	97.2
Female	0.951 ^a	0.0237	0.894-0.983	98.4	87.1	89.6	97.9
Male	0.968 ^a	0.0144	0.917-0.992	93.4	92.6	93.4	92.6
<i>Without protocol</i>							
Total	0.973 ^a	0.0086	0.942-0.991	88.4	95.8	94.4	91.3
Female	0.984 ^a	0.0100	0.935-0.999	94.1	96.9	94.1	97.0
Male	0.986 ^a	0.0077	0.939-0.999	97.6	92.4	86.8	98.4
BIA 4							
<i>With protocol</i>							
Total	0.932 ^a	0.0156	0.890-0.962	83.2	88.3	84.9	86.9
Female	0.907 ^a	0.0261	0.838-0.953	88.5	75.9	80.6	85.4
Male	0.970 ^a	0.0134	0.914-0.994	97.1	81.2	73.3	98.2
<i>Without protocol</i>							
Total	0.929 ^a	0.0161	0.886-0.959	85.3	85.9	82.7	88.0
Female	0.899 ^a	0.0277	0.829-0.947	75.4	88.9	88.5	76.2
Male	0.980 ^a	0.0106	0.930-0.998	91.2	95.5	91.2	95.5

AUC, area under the curve; BIA 1, biodynamics horizontal tetrapolar device, model 450; BIA 2, tanita vertical tetrapolar device, model BC-558; BIA 3, biospace equipped with eight tactile electrodes, model InBody 230; BIA 4, tanita vertical bipolar device, model 2220; CI, confidence interval; NPV, negative predictive value; PPV, positive predictive value; SD, standard deviation; Sensit, Sensitivity; Specif, specificity.

ROC curve: BIA vs. DXA.

^a $p < 0.001$.

For males, however, BIA 2 was not shown to be an accurate option.

The results showed that the protocol did not influence results in female adolescents, which, together with the similarity to DXA, demonstrates that the assessment without protocol could be used for females using any of the devices. For males, although devices 1 and 4 were influenced by the protocol, both assessments were similar to that of DXA. Thus, it is suggested that when using these devices, one of the two forms of assessment should be standardized.

The protocol used aimed to standardize factors that may influence BIA assessment accuracy, mainly related to the

state of hydration, such as beverage consumption, menstrual period, and physical activity. Guidelines related to the technical aspects of the devices, which were provided by the manufacturers, were also observed.¹⁶

BIA is based on the passage a low-intensity electric current through the body of the individual; impedance, resistance, reactance, and phase angle values are determined, through which body composition is estimated. These values are strongly related to body hydration, as water is a good conductor of electricity, while fat is not. If the tissues are in atypical conditions of hydration, the method accuracy is compromised.^{10,27}

The protocol, however, can compromise adherence of adolescents in population studies, as it requires effort and interest to follow the established requirements. It was observed, however, that some of the analyzed devices, particularly BIA 3, had good results even without the proposed standardization, suggesting an alternative when it is not possible to conduct the protocol.

There have been no studies in the literature that verified the influence of the protocol on the use of BIA in adolescents. However, some studies investigated differences between assessments after the consumption of food, one of the items that comprised the protocol used in the present study.

Vilaça et al.,²⁸ when evaluating 41 elderly Brazilian males, used data obtained by tetrapolar BIA and compared to DXA, after fasting and after eating a meal. No differences were observed between the measurements ($p > 0.05$). These results corroborate those found in the present study, confirming the usefulness of the assessment without protocol.

Conversely, Gallagher et al.,¹¹ when studying the influence of meals with different compositions on the results estimated by BIA in 28 Australian adults of both genders, reported that there was a significant variation in impedance and consequently, on BF estimate, after consumption of meals. Similar results were demonstrated by Slinde and Rossander-Hulthen,¹⁰ when they evaluated healthy adults by tetrapolar BIA method, before and after the consumption of three standardized meals during a 24-hour period. The authors concluded that the impedance measurement decreased approximately two to four hours after a meal ($p < 0.05$), causing variation of up to 8.8% (women) and 9.9% (men) in BF%, underestimating it.

The two studies presented opposite results to those of the present study, showing the influence of the protocol on the assessments, but because they were not compared with DXA, it is impossible to know whether the assessment after consumption of meals would also be helpful.

Regarding the agreement of assessments with BIA and DXA, BIA had three with better results, and although none of the devices had a kappa index > 0.8 (almost perfect agreement)¹⁷, when analyzed together with the other results, it was observed that the results confirmed the possibility of using the assessment without protocol.

The ROC curve analysis showed again the usefulness of BIA in the absence of a protocol. There was no difference between areas with and without protocol for any of the devices, indicating the capacity of this assessment in predicting BF% increase, as all constructed curves were significant ($p < 0.001$).

BIA 3 showed the greatest areas for the general and stratified population. It is observed that the general population and the female gender had higher sensitivity when adopting the protocol, which demonstrates its capacity to detect a greater number of adolescents with excess BF. As for the male gender, the highest sensitivity was demonstrated at the evaluation without the protocol, which again highlights its usefulness.

This device has a tetrapolar system, which differs from the others, since it has by eight tactile electrodes and is multifrequency. The combination of these factors appears to ensure more sensitivity when estimating body composition in

adolescents, while the protocol did not influence the results in any of the analyzed situations.

The other BIA devices, which are of lower cost and more available for health services and that also showed moderate sensitivity, specificity, and positive and negative predictive values, can be used with caution at the population level, in the absence of more sensitive methods.

The assessment of body composition of any adolescent performed by methods that are not considered to be the "gold standard" should be made considering the possible errors and should not annul the importance and the need for prevention activities and/or control of excess BF, whatever the results.

Conclusion

Based on the results, it was concluded that electrical bioimpedance has good predictive capacity to estimate excess BF in adolescents, and that when it is not possible to perform the protocol, the results are also similar to those of DXA, thus allowing its use in population studies.

The tetrapolar device with eight tactile electrodes showed the highest sensitivity and the best results for the overall population and for the female gender, with protocol, and for the male gender, without protocol.

A high prevalence of adolescents with high percentage of BF was observed, which suggests the importance of specific health care programs in this population, aiming to correct dystrophies and prevent cardiovascular and metabolic disorders in adulthood.

Due to the widespread use of electric bioimpedance devices, studies with other age groups, in the presence and absence of a protocol, are required to confirm its importance and indicate the reliability of the results for the entire population.

Funding

Financial support was received from the following agencies: Fundação de Amparo à Pesquisa do Estado de Minas Gerais (FAPEMIG)/Process No. APQ-01618-10 and Conselho Nacional de Desenvolvimento Científico e Tecnológico (CNPq)/Process No. 485986/2011-6.

Conflicts of interest

The authors declare no conflicts of interest.

References

1. Organização Mundial da Saúde - OMS. Nutrition in adolescence - issues and challenges for the health sector WHO. Geneva: WHO; 2005.
2. Instituto Brasileiro de Geografia e Estatística - IBGE. Pesquisa de orçamentos familiares 2008-2009: despesas, rendimentos e condições de vida. Rio de Janeiro; 2010.
3. Kim HA, Lee Y, Kwon HS, Lee SH, Jung MH, Han K, et al. Gender differences in the association of insulin resistance with metabolic risk factors among Korean adolescents: Korea National Health and Nutrition Examination Survey 2008-2010. *Diabetes Res Clin Pract.* 2013;99:54-62.

4. Jaeger AS, Barón MA. Uso de la bioimpedancia eléctrica para la estimación de la composición corporal en niños y adolescentes. *An Venez Nutr.* 2009;22:105.
5. Kim H, Kim CH, Kim DW, Park M, Park HS, Min SS, et al. External cross-validation of bioelectrical impedance analysis for the assessment of body composition in Korean adults. *Nutr Res Pract.* 2011;5:246–52.
6. Siervogel RM, Demerath EW, Schubert C, Remsburg KE, Chumlea WC, Sun S, et al. Puberty and body composition. *Horm Res.* 2003;60:36–45.
7. Linares CL, Ciangura C, Bouillot J-L, Coupaye M, Declèves X, Poitou C, et al. Validity of leg-to-leg bioelectrical impedance analysis to estimate body fat in obesity. *Obes Surg.* 2011;21:917.
8. Rodrigues MN, Silva SC, Monteiro WD, Farinatti PTV. Estimativa da gordura corporal através de equipamentos de bioimpedância, dobras cutâneas e pesagem hidrostática. *Rev Bras Med Esporte.* 2001;7:125–31.
9. Eickemberg M, Oliveira CC, Roriz AKC, Sampaio LR. Bioimpedância elétrica e sua aplicação em avaliação nutricional. *Rev Nutr.* 2011;24:873–82.
10. Slinde F, Rossander-Hulthen L. Bioelectrical impedance: effect of 3 identical meals on diurnal impedance variation and calculation of body composition. *Am J Clin Nutr.* 2001;74:474–8.
11. Gallagher M, Walker KZ, O’Dea K. The influence of a breakfast meal on the assessment of body composition using bioelectrical impedance. *Eur J Clin Nutr.* 1998;52:94–7.
12. Brasil – Ministério da Saúde – DATASUS Dados epidemiológicos Viçosa 2010. [cited 03 Mar]. Available from: <http://tabnet.datasus.gov.br/cgi/deftohtm.exe?ibge/cnv/popmg.def>.
13. Cardoso Chaves O, do Carmo Castro Franceschini S, Machado Rocha Ribeiro S, Ferreira Rocha Sant’Ana L, Garçon de Faria C, Eloiza Priore S. Comparison of the biochemical, anthropometric and body composition variables between adolescents from 10 to 13 years old and their parents. *Nutr Hosp.* 2012;27:1127–33.
14. de Onis M, Onyango AW, Borghi E, Siyam A, Nishida C, Siekmann J. Development of a WHO growth reference for school-aged children and adolescents. *Bull World Health Organ.* 2007;85:660–7.
15. Lohman TG. Assessing fat distribution In: *Advances in body composition assessment: current issues in exercise science.* Illinois: Human Kinetics. Champaign. 1992:57–63.
16. Barbosa KBF. Consumo alimentar e marcadores de risco para a síndrome metabólica em adolescentes do sexo feminino: comparação entre instrumentos de inquérito dietético [Dissertation]. Viçosa MG: Universidade Federal de Viçosa; 2006.
17. Landis JR, Koch GG. The measurement of observer agreement for categorical data. *Biometrics.* 1977;33:159–74.
18. Martinez EZ, Louzada-Neto F, Pereira BB. A curva ROC para testes diagnósticos. *Caderno de Saúde Coletiva.* 2003;11:7–31.
19. Coelho LG, Cândido AP, Machado-Coelho GL, de Freitas SN. Association between nutritional status, food habits and physical activity level in schoolchildren. *J Pediatr (Rio J).* 2012;88:406–12.
20. Faria ER, Franceschini Sdo C, Peluzio Mdo C, Sant’Ana LF, Priore SE. Correlação entre variáveis de composição corporal e metabólica em adolescentes do sexo feminino. *Arq Bras Cardiol.* 2009;93:119–27.
21. Serrano HM, Carvalho GQ, Pereira PF, Peluzio Mdo C, Franceschini Sdo C, Priore SE. Composição corpórea, alterações bioquímicas e clínicas de adolescentes com excesso de adiposidade. *Arq Bras Cardiol.* 2010;95:464–72.
22. Carvalho GQ, Pereira PF, Serrano HM, Franceschini Sdo C, de Paula SO, Priore SE, et al. Peripheral expression of inflammatory markers in overweight female adolescents and euthropic female adolescents with a high percentage of body fat. *Appl Physiol Nutr Metab.* 2010;35:1–7.
23. Pereira PF, Serrano HM, Carvalho GQ, Lamounier JA, Peluzio Mdo C, Franceschini Sdo C, et al. Body fat location and cardiovascular disease risk factors in overweight female adolescents and euthropic female adolescents with a high percentage of body fat. *Cardiol Young.* 2012;22:162–9.
24. Vieira PR, Faria E, Faria F, Sperandio N, Araújo C, Stofeles R, et al. Fatores associados à adiposidade em adolescentes do sexo feminino eutróficas com adequado e elevado percentual de gordura corporal: elaboração de um modelo de risco. *Arch Latinoam Nutr.* 2011;61:279–87.
25. Gonçalves VSS, Chaves OC, Ribeiro SMR, Sant’ana LF, Franceschini Sdo C, Priore SE. Disponibilidade domiciliar de lipídeos para consumo e sua relação com os lipídeos séricos de adolescentes. *Rev Paul Pediatr.* 2012;30:229–36.
26. Costa RF, Santos NS, Goldraich NP, Barski TF, de Andrade KS, Krueh LF. Metabolic syndrome in obese adolescents: a comparison of three different diagnostic criteria. *J Pediatr (Rio J).* 2012;88:303–9.
27. National Institutes of Health Technology Assessment Conference Statement, 12. Nutrition.: Bioelectrical impedance analysis in body composition measurement.; 1994. p. 1–35.
28. Vilaca KH, Ferriolli E, Lima NK, Paula FJ, Moriguti JC. Effect of fluid and food intake on the body composition evaluation of elderly persons. *J Nutr Health Aging.* 2009;13:183–6.